

**SUNY Downstate Medical Center -University Hospital
of Brooklyn Network
Department of Pathology Policy and Procedure**



Subject: LAB 6 TEST PERFORMANCE ASSESSMENT

Prepared By: Alix R Laquerre

LTR: LTR12807

Edit Approved By: [Howanitz MD, Peter \(Electronic Signature Timestamp: 4/18/2013 3:46:00 PM\)](#)

Reviewed By: [Howanitz MD, Peter \(Electronic Signature Timestamp: 4/18/2013 3:46:00 PM\)](#)

[Laquerre MS, Alix \(Electronic Signature Timestamp: 4/18/2013 11:24:37 AM\)](#)

[Laquerre MS, Alix \(Electronic Signature Timestamp: 4/18/2013 11:24:37 AM\)](#)

Supporting Documents:

Approval Workgroup: Laboratory Administration
Approval Group

Revision: 3

PURPOSE: To determine accuracy and reliability of analytic results on patient samples for which no external proficiency testing is offered.

POLICY: Laboratory analyses for which no external proficiency is offered will be assessed at least semi annually for accuracy and reliability by performing split sampling with a reference laboratory and/or inter-laboratory split sampling.

Criteria for unacceptable results will be defined in each laboratory section. Results will be documented and reviewed. Action taken for unacceptable results will be documented and presented at the Departmental P.I. meeting.

Results will be documented and reviewed. Action taken for unacceptable results will be documented.

Test with no external proficiency testing:

HEMATOLOGY

Urea Solubility

Heinz Body Stain

LAP

LE Latex